

DEPARTMENT OF**X-RAY EQUIPMENT SAFETY CHECKLIST****INSTRUMENT:****LOCATION:****CARRIED OUT BY (print name):****SIGNATURE:**

LIGHTS	COLOUR	OK – CHECKED (v)
GENERATOR ON		
HT/X-RAYS ON		
GENERATOR		
REMOTE		
SHUTTER WARNING		
SHUTTER LAMP OPEN		
CLOSED		
INTERLOCK OPEN		
CLOSED		
SAFETY CIRCUIT		
CORRECT		
FAILURE		
WATER SAFETY		

INTERLOCKS	COMMENTS	OK – CHECKED (v)
GENERATOR KEY SWITCH		
DOOR INTERLOCKS		
FRONT		
BACK		
SIDE		
PANEL INTERLOCKS	(No.)	
CAMERA/INSTRUMENT/SLIT SYSTEM INTERLOCKS	(No.)	
OVERRIDE KEY FUNCTION CORRECT		
POSITION OF OVERRIDE KEY(S)		
WATER SAFETY		

ADDITIONAL NOTES: