

The Numbers and Validation Games

Mark Wills, Department of Medicine BSO
Chair University Biological safety Sub-committee



The Numbers (Terminology) game -Example extract from Risk Assessments




Hazards identified: *n.b.* if your assessment includes substances hazardous to health use COSHH form.

Summary of work: The work includes in vitro infection of cells with influenza A X31 virus. All the reagents and solutions are prepared under sterile conditions in a hood, in cat 2 area.



Hazards identified and control measures to reduce the level of risk:

- (1) Influenza A X-31 virus has been prepared at the [REDACTED], where it is being routinely used for in vivo and in vitro infections.
- (2) To ensure sterile preparation of the reagents preparation and dilution of working solutions will be prepared under laminar flow in sterile conditions.



Examples from Risk Assessments/Accidents/Incident reports

In a hood

We bombed
The hood

In the
cabinet

Class II
pathogen




The Cat 2
lab

Bug plates

The Class 2
Tissue culture room

Why is this important? “You know what I mean, you’re being a pedant!”

- Accurate and efficient communication
- Risk assessments or reports with incorrect use of terminology might suggest a poor understanding by those writing them
- Auditing/assessment by outside agencies (HSE, DEFRA etc) – correct use of terminology is an expectation of an institution that understands what it is doing.
- Professionalism



Microbiological Safety Cabinets Class I, II, III



Containment level (CL) - The Laboratory
Hazard group (HG) – The Pathogen
Activity class(Class 1) – Genetic Modification

Containment level 1 (CL 1)
for **hazard group 1 (HG 1)**
and **activity class 1 (Class 1)**.

Containment level 2 (CL 2)
for **hazard group 2 (HG 2)**
and **activity class 2 (Class 2)**.

Containment level 3 (CL 3)
for **hazard group 3 (HG 3)**
and **activity class 3 (Class 3)**

**Specified Animal Pathogens
(SAPO) Groups 1,2,3 and 4**

In addition to the usual risk assessments for Genetic Modification (GM), Control of Substances Hazardous to Health (COSHH), Specified Animal Pathogens Order (SAPO) etc, departments must consider if the pathogen is one listed in Schedule 5.

Eg A hazard group 3 pathogen, or a GM project assessed as class 3 clearly has to be used in a containment level (CL) 3 facility; but if the pathogen is listed in Schedule 5, stringent physical (and other) security measures over and above those required for CL3 will be required.

Categories A, B and C

The Validation Game

Autoclaves

- Certificates
- Thermocouple checks
- Biological monitoring
- Autoclave tape!!

Microbiological Safety Cabinets

- In flow and down flow air speeds
- Visual Inspection and operational control
- KI-DISCUS
- Real in use setup
 - Room air flows – air-conditioning units
 - in cabinet equipment
- How do you test the alarms?

Gas monitors

- alarm tests
- calibration

Temperature monitoring critical equipment

+4, -20, -80 °C etc

eg TSCAN monitoring solution

How do you test it works ? Required ?

Magnehelic Pressure gauges

calibrated - who when?

low pressure alarms – how do you test?

Legionella monitoring

Risk assessment

By who?

Of what?

What do the results mean?

Autoclaves – Validation, Calibration and Monitoring

Biological Safety Series

April 2019

Autoclaves: Validation and Monitoring

Occupational Health and
Safety Service
HSD164B (rev 2)

Biological Safety Series

April 2019

Waste Inactivation Validation

Occupational Health and
Safety Service
HSD173B (rev 1)



Validation: Establishing documented evidence that a disinfection process will consistently inactivate target organisms under defined conditions of use.

- *Thermocouple 'mapping' should be used. This involves placing multiple (usually 12+) independent thermocouples at various sites (including the most inaccessible) within both 'typical' and 'difficult to penetrate' simulated loads. **ANNUAL***

Monitoring: To observe or record the activity or performance of a device.


- *"Cycle run report" –has it passed or failed – **EVERY RUN***
- *Biological indicators – spore strips and culture – **MONTHLY?***

Calibration: The act of checking or adjusting (by comparison with a standard) the accuracy of a measuring instrument.


- *Does the autoclave measure the correct temperature and pressure? **ANNUAL***

Pressure Systems Safety Regulations (PSSR) 2000

Autoclaves (ALL Pressure vessels) need to be (**annually**) inspected by a representative for the University's insurers to comply. *Bureau Veritas (BV) acts as the Competent Person to carry out this inspection. All autoclaves will automatically be registered onto the BV SWIFT database to which the departmental lead contacts have access.*




Testing and validation of microbiological safety cabinets



There is a statutory requirement to carry out testing (every 6 months or 12-14 months depends on Containment Level) and validation of microbiological safety cabinets. Performance tests are carried out to BS EN12469 and include

- HEPA filter integrity
- airflow velocity and visualization tests,
- operation of airflow indicators & alarms
- KI-Discus (aperture protection factor) tests.

In Use Testing of microbiological safety cabinets



Is work usually carried out in an empty cabinet?

- Equipment that is **USUALLY** in the MSC should be in the MSC when it is tested.
- If a KI test is being performed **BAG** and protect the equipment!!

ESCO.

WORLD CLASS. WORLDWIDE.

BIOSAFETY CABINET FIELD CERTIFICATION

Manufacturer: ESCO Class: II
 Model: ACE 2498 Serial No. 2017 123015
 Location: Cambridge Test Report No. 13709
 Test Date: 23-03-18 Due Date: 22-03-19
 NSF 49 ☐ EN 12469 ☒ Other ☐

Test performed to manufacture's specification, NSF 49 or other international guideline	Pass	Fail	Not Tested
Primary Test			
Inflow Velocity	☒	☐	☐
Downflow Velocity	☒	☐	☐
Airflow Smoke Patterns	☒	☐	☐
Supply HEPA Filter Leakage	☒	☐	☐
Exhaust HEPA Filter Leakage	☒	☐	☐
Site Installation Assessment	☐	☐	☐
Secondary Test (Optional Only)			
KI-Discus Test (Operator Protection)	☒	☐	☐
Light Intensity	☐	☐	☐
UV Intensity	☐	☐	☒
Noise Level	☒	☐	☐
Certification Result	☒	☐	☐

Remarks: _____

Certified By: Thavakumar Certificate No.: _____

Signature: T. D. Email: sgp.svc@escoglobal.com
sgp.sal@escoglobal.com

Escoc Micro Pte Ltd

21 Changi South Street 1 Singapore 486777

Tel: +65 65420833 Fax: +65 65426920 Website: www.escoglobal.com

Escoc Gb Ltd	Tel: +44-0-1226361529	Fax: +44-0-1226 741709
Escoc Technologies, Inc.	Tel: +1-215 441 9661	Fax: +1-215 441 9660
Escoc Philipines, Inc.	Tel: +632-8281527	Fax: +632-8281527
Escoc Micro (M) Sdn. Bhd.	Tel: +603-5634 5833	Fax: +603-56387895
Escoc Vietnam Company Ltd.,	Tel: +844-62691460	Fax: +844-62691461

**ASSET REGISTER
(DO NOT REMOVE)**

112003

**UNIVERSITY OF CAMBRIDGE
CLINICAL SCHOOL**

V A L I D A T I O N L O G	 ch_ts Crowthorne Hi-Tec Services		Beech House, Ancells Business Park, Ancells Road, Fleet, Hampshire GU51 2UN Tel: +44 (0)1252 372333 • Fax: +44 (0)1252 370060 email: service@chts.co.uk www.crowthornehitec.co.uk		
			Other offices:		
	Scotland: 16 Flakfield, College Milton North, East Kilbride, Glasgow, G74 1PF Telephone: +44 (0)1355 228030 Fax: +44 (0)1355 264744 email: service@chts.co.uk www.crowthornehitec.co.uk		Ireland: Unit 4C Bymac Centre, North West Business Park, Bullycoolin, Blanchardstown, Dublin 15 Telephone: +353 (0)1 8243670 Fax: +353 (0)1 8243671 email: sales@chtai.com www.crowthornehitec.co.uk		
	MAKE	MODEL	SER No	I.D. No	
	ESCO	CLASS II (R)	201-123015	—	
	THIS EQUIPMENT HAS BEEN INSPECTED AND VALIDATED BY CROWTHORNE HI-TEC ANY PROBLEM FOUND WITH THIS UNIT PLEASE CONTACT 01252 372333 OR SERVICE@CHTS.CO.UK				
VALIDATION DATE	RESULT PASS / FAIL	K.I. TEST PASS / FAIL	TEST SHEET No.	NEXT VALIDATION	ENGINEER
15/11/18	PASS	—	94660	MAY19	JA.
10/5/19	PASS	PASS	1568	NOV19	JA



Environmental Validation Solutions Ltd

67 Hoylake Crescent Ickenham Middlesex UB10 8JF

Tel: 020 8335 3532

E-mail: service@evs-uk.com

Test Report
Certificate No.

L832902

Customer Cambridge university
Building Dept of medicine Room HR08 05 800 Cat
Cabinet Manufacturer CAS Type CLII Tripass
Serial No C1423-2 Site No 0 EVS Ref

Site Contact A.Brownlee Job No L8329
Lab Contact A.Brownlee Engineer S.Palmer
Previous Certificate n/a Date 12-Nov-13
Phone no Order no 0

Test Equipment

Anemometer TA 460 Photometer JM8000
Serial No TA4600944002 Serial No JM8-8028
Calibration Date 31/04/2013 Calibration Date 22-Nov-12

Cabinet Function and Examination

Alarm Audible Working Visual Working
Fan Exhaust Working Downflow Working
Fan Run On n/a min Airflow Meter Safe
Duct Pressure Open n/a Pa Closed n/a Pa
Sound ok Hour Meter n/a hrs
Indicator Switches Working Funigation n/a
Gas Solenoid n/a Socket Test Working
Lighting Working UV Lamp n/a

Cabinet Condition

Exterior Surface ok Work Surface ok
Interior Surfaces ok Viewing Panel ok
Duct Connection ok Closure Panels ok

Comments

If cabinet goes into alarm it will still have the containment protection factor as it is set up with the setting parameter high enough so as to comply if the room negative pressure drops to alarm levels.

Particle Counts Particles per Cu m

Position	≥ 0.3µm	≥ 0.5µm	≥ 5.0µm
1			
2			
3			
4			
5			
6			

Counter
Serial No
Calibrated

Passes to GMP Class A Yes/No

Air Flow Measurements m/s

Inflow Average Velocity 0.61 m/s
Pass

0.63 m/s 0.56 m/s 0.64 m/s

Acceptance Criteria : Inflow to be not < 0.4 m/s

Downflow Average Velocity 0.41 m/s
Pass

0.40 m/s 0.38 m/s 0.38 m/s 0.41 m/s

0.41 m/s 0.40 m/s 0.46 m/s 0.44 m/s

Acceptance Criteria : Downflow to be 0.35 to 0.5 m/s ± 20%

Filter Tests

Down-flow/Out-flow filter	1st Exhaust / Main filter	2nd Exhaust / In-flow filter	Thimble / Room filter	Other filter
Serial No	Serial No	Serial No	Serial No	
Penetration	Penetration	Penetration	Penetration	
Media <0.003 %	Media <0.003 %	Media <0.003 %	Media n/a %	
Seals <0.003 %	Seals <0.003 %	Seals <0.003 %	Seals n/a %	
Pass	Pass	Pass		

Acceptance Criteria : Penetration to be < 0.003 %

KI-Discus Aperture Protection Factor Results

No	X	OP	Y	OP	XI	OP	YI	OP
M1	5	12.4	9	6.9	6	10.3	14	4.4
M2	8	7.8	5	12.4	6	10.3	8	7.8
M3	6	10.3	7	8.9	8	7.8	10	6.2
M4	2	31.0	10	6.2	6	10.3	7	8.9
M5	6	10.3	4	15.5	12	5.2	11	5.6

L1								
L2								
L3								
L4								
L5								

R1								
R2								
R3								
R4								
R5								

These Aperture Protection Results DO comply with the requirements of EN12469:2000

K.I.Discus serial No K12/02-0163

Calibration 19-Dec-2012

Acceptance Criteria : Aperture Protection to be not less than 1×10^{-5}

The above test data was correct at the time of the tests. If the equipment is moved, modified or other environmental factors change the equipment may require a re-test

At the time of test this Cabinet

Passes

E.S.R. Number

Engineers Signature

Customer Signature

Date

Next Test Due

Label Attached

May-14

yes

Gas monitors – what sort oxygen depletion or active detection of gas of interest, maybe BOTH.

- How do you alarm test them?
- How often are they tests performed?
- How do you have them calibrated?
- Who is responsible?
- Record the monitoring/calibration
- Installed systems and personal monitors



Chemical Safety Guidance

July 2018

Compressed Gas User Guidance

Occupational Health and Safety Service
HSD032C (rev 5)



Magnehelic Pressure gauges
calibrated - who and when?
low pressure alarms – how do you test?

Temperature monitoring critical equipment



- Could be +4oC, -20oC, -60-80oC
- Critically important unique samples
- Bio banks – Human Tissue – HTA
- Automatic monitoring, remote reporting and control
- Alarm cascades by mobile phone, text message and email

NB other suppliers of monitoring services are available!!



Legionella monitoring and data logging

Biological Safety Series

June 2017

Legionella Awareness

Frequently Asked
Questions and what you
can do to help

Occupational Health and Safety Service
HSD175B (rev 1)


University or embedded
accommodation?

Risk assessment
By who?

Of what

- Ice Machines
- Plumbed eyewash
- Sinks

What do the results
mean?



The Validation Game: Check, Calibrate and Cascade

Closing the loop

- Facilities have had the equipment serviced
 - Where do the reports go?
 - What do the reports mean?
 - Actions to be taken?
 - Who needs to know the result and how is the information passed on?

Lay interpretation of engineers reports

HSE CL3 inspection question – How do you know a user understands the results of the inspection and the written report?

- Verbal report from engineer to user/facilities before leaving
 - Did it pass
 - Any concerns going forward
 - Close the loop!

Permits to work

- NOT just relevant to higher CL3 containment also CL2
- What information was the service engineer given?
- What was disinfected?
- What they can and cannot touch?
- What PPE should they wear?
- What should they do if the fire alarms go off?

The Validation Game: Check, Calibrate and Cascade

Record keeping - keeping evidence and being able to provide it for audit /inspection to HSE and others

- Permits and service reports
- Maintenance inspection schedules
 - Who is responsible
 - Who keeps the schedules
 - How do the users know service is required (before it becomes an issue)
 - Autoclaves annually under the written scheme (after that date CAN NOT BE USED)
 - *Bureau Veritas* SWIFT database
- A clear, formalized check list system to manage the schedule of inspections, servicing etc.
- A clear indication of who is responsible and who is responsible if that person is not available.
- A schedule to check the schedule – Seriously!! Comment from HSE inspection
- RECORD, RECORD, RECORD – Time, Date and Sign