The Numbers and Validation Games Mark Wills, Department of Medicine BSO Chair University Biological safety Sub-committee











The Numbers (Terminology) game -Example extract from Risk Assessments



Hazards identified: *n.b.* if your assessment includes substances hazardous to health use COSHH form.


Summary of work: The work includes in vitro infection of cells with influenza A X31 virus. All the reagents and solutions are prepared under sterile conditions in a hood, in cat 2 area.






Hazards identified and control measures to reduce the level of risk:


- (1) Influenza A X-31 virus has been prepared at the where it is being routinely used for in vivo and in vitro infections.
- (2) To ensure sterile preparation of the reagents preparation and dilution of working solutions will be prepared under laminar flow in sterile conditions.



Examples from Risk Assessments/Accidents/Incident reports








Why is this important? "You know what I mean, you're being a pedant!"

- Accurate and efficient communication
- Risk assessments or reports with incorrect use of terminology might suggest a poor understanding by those writing them
- Auditing/assessment by outside agencies (HSE, DEFRA etc) –
 correct use of terminology is an expectation of an
 institution that understands what it is doing.
- Professionalism



Biological Safety Series

Microbiological Safety Cabinets Class I, II, III



Containment level (CL) - The Laboratory
Hazard group (HG) – The Pathogen
Activity class(Class 1) – Genetic Modification

Containment level 1 (CL 1) for **hazard group 1** (HG 1) and **activity class 1** (Class 1).

Containment level 2 (CL 2) for **hazard group 2** (HG 2) and **activity class 2** (Class 2).

Containment level 3 (CL 3) for hazard group 3 (HG 3) and activity class 3 (Class 3)

Specified Animal Pathogens (SAPO) Groups 1,2,3 and 4

In addition to the usual risk assessments for Genetic Modification (GM), Control of Substances Hazardous to Health (COSHH), Specified Animal Pathogens Order (SAPO) etc, departments must consider if the pathogen is one listed in Schedule 5.

Eg A hazard group 3 pathogen, or a GM project assessed as class 3 clearly has to be used in a containment level (CL) 3 facility; but if the pathogen is listed in Schedule 5, stringent physical (and other) security measures over and above those required for CL3 will be required.

Categories A, B and C

Why is this important? Example

								EORI No: G	B 125 5067 30 065
Supplier Number	Supplier VAT Number	Customer Account No.	Payment Terms Supp		Supplier Contact Phone		ie	Supplier Contact Fax	
2054012	GB596360800	1088	30 Days		08005830040			01865772482	
Line Part Number/Description No.			Delivery Date	Qty	UOM	Unit Price GBP	Total GBP		VAT Reference
1 Cat III dry ice delivery from Cam	bridge to		Needed: 12-MAY-2022	1	Each	204.75		204.75	STANDARD
STANDARD = Standard Rate VAT						Total Excl. VAT:		204.75	

This is an order request raised by another University so that a HG3 pathogen could be sent from Cambridge to them. The other University selected the courier company to be used.

Q How confident would you be about this?

Why is this important? Example

Supplier Number	Supplier VAT Number	Customer Account No.	Payment Terms Suppl			r Contact Phon	e Supplier C	Contact Fax
2054012	GB596360800	1088	30 Days		08005830	0040	0186577248	32
Line Part Number/Description			Delivery Date	Qty	UOM	Unit Price GBP	Total GBP	VAT Reference
Cat III dry ice delivery from Cam	bridge to		Needed: 12-MAY-2022	1	Each	204.75	204.75	STANDARD
TANDARS = Standard Rate VAT			1	I	1	Total Excl. VAT:	204.75	

This is an order rquest raised by another University so that a HG3 pathogen could be sent from Cambridge to them. The other University selected the courier company to be used.

Q. How confident would you be about this?

A. It should be a **RED** flag

The Validation Game

Autoclaves

Certificates

Theromocouple checks

Biological monitoring

Autoclave tape!!

Microbiological Safety Cabinets

In flow and down flow air speeds

Visual Inspection and operational control

KI-DISCUS

Real in use setup

- Room air flows air-conditioning units
- in cabinet equipment

How do you test the alarms?

Gas monitors

alarm tests calibration

Temperature monitoring critical equipment

+4, -20, -80 oC etc

eg TSCAN monitoring solution

How do you test it works? Required?

Magnehelic Pressure gauges

calibrated - who when?

low pressure alarms – how do you test?

Legionella monitoring

Risk assessment

By who?

Of what?

What do the results mean?

Autoclaves – Validation, Calibration and Monitoring

Biological Safety Series

April 2019

Autoclaves:

Validation and Monitoring

Occupational Health and Safety Service HSD164B (rev 2)

Pressure Systems Safety Regulations (PSSR) 2000

Autoclaves (ALL Pressure vessels) need to be (**annually**) inspected by a representative for the University's insurers to comply. *Bureau Veritas (BV) acts as the Competent Person to carry out this inspection. All autoclaves will automatically be registered onto the BV SWIFT database to which the departmental lead contacts have access.*










Testing and validation of microbiological safety cabinets





There is a statutory requirement to carry out testing (every 6 months or 12-14 months depends on Containment Level) and validation of microbiological safety cabinets. Performance tests are carried out to BS EN12469 and include

- HEPA filter integrity
- airflow velocity and visualization tests,
- operation of airflow indicators & alarms
- KI-Discus (aperture protection factor) tests.

Is work usually carried out in an empty cabinet?

- Equipment that is USUALLY in the MSC should be in the MSC when it is tested.
- If a KI test is being performed BAG and protect the equipment!!

EVS Enviro	lidation Solutio	tion Solutions Ltd 67 Hoylake Crescent Ickenham Middlesex UB10 8JF Tel: 020 8335 3532 E-mail: service@evs-uk.com					Test Report Certificate: No. L832902				
Customer Cambridge university			Site Contact A.Brownlee Job No			ii ii	L8329	Test Equipment			
Building Dept of medicine Room HR08 05 800			Lab Contact	A.Brownlee	Engineer	S	S.Palmer Anemometer: TA 460 Photometer:				
Cabinet Manufacturer CAS Type CL			Previous Certificate	n/a	Date	12	-Nov-13	Anemometer			
Serial No C1423-2 Site	FVSRef						Calibration Date: 31/04/2013 Calibration Date: 22-Nov-12				
Cabinet Function and Examination			Air Flow Measurem	ients m/s			KI-	Discus Aperture Prot	Discus Aperture Protection Factor Results		
Alam Audible Working Visual	Working	Inflow	Avera	ige Velocity 0	.61 m/s	No:	X OP	YOP	X1 OP Y1	OP	
Fan Exhaust Working Downflow	Working			Pass		M1:	5 12.4	9 6.9	6 10.3 14	4.4	
Fan Run On n/a min Airflow Me							8 7.8	5 12.4	6 10.3 8	7.8	
Duct Pressure Open n/a Pa Closed	n/a P						6 10.3	7 8.9	8 7.8 10	6.2	
Sound ok Hour Meter	()()()()	;		· · · · · · · · · · · · · · · · · · ·	•	2000	2 31.0	10 6.2	6 10.3 7	8.9	
Indicator Switches Working Fumigation Gas Solenoid n/a Socket Tes	n/a Working	0.63	0.56	0.64		M5	6 10.3	4 15.5	12 5.2 11	5.6	
Lighting Working UV Lamp	working n/a	0.63 0.56 0.64 m/s m/s m/s			:EE:						
Cabinet Condition		1 100	1 111/5	1 11/5	<u>.</u>	1.2					
Exterior Surface ok Work Surfa	ce;∵; ok					L3					
Interior Surfaces ok Viewing Pa	.,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,					L4		· 			
Duct Connection ok Closure Panels ok		Acceptance Criteria: Inflow to be not < 0.4 m/s						· 			
		Pownflow Average Velocity 0.41 m/s							····		
Comments	The Section 1			Pass		RI					
If cabinet goes into alarm it will still have the protection factor as it is set up with the settin high enough so as to comply if the room negat drops to alarm levels.	0.40 m/s	0.38 0 m/s	.38 0.41	m/s	R2 R3 R4 R5			h			
		7	,	, , , , , , , , , , , , , , , ,	;	These Ape	erture Protection	Results DO comply wi	th the requirements of EN12469:	2000	
		0.41	0.40 0	16		77. Y	D' '137	7710/00 01/0	G (1)		
***************************************		0.41 0.40 0.46 0.44 K					K.I Discus serial No K12/02-0163 Calibration 19-Dec-2012 Acceptance Criteria: Aperture Protection to be not less than 1 x 10 5				
Particle Counts Paticles per Cu n		·/			The above			the equipment is moved, modified			
Position $\geq 0.3 \mu \text{m}$ $\geq 0.5 \mu \text{m}$	≥ 5.0µm	Accenta	nee Critoria - Downflow to be 0	25 to 0 5 m/s 1/ 200/				change the equipment may		4-;	
1		Filter Tests						At the time of test this Ca	ibinet. Passes		
2		Down-flow/Out-flow filter 1st Exhaust / Main filter 2nd Exhaust / In-flow filter Thimble / Room filter Other filter				Other filter	E.S.R. Number				
3		Serial No	Serial No	Serial No	Serial No						
4								Engineers Signature			
		Penetration	Penetration P	enetration	Penetration				·		
		Media <0.003 %	Media <0.003 %	Media <0.003 %	Media	n/a %		Customer Signature			
Counter									}		
		Seals <0.003 %	Seals <0.003 %	Seals <0.003 %	Seals	n/a %		Date			
	Pass	Pass	Page				Novt Take Date	Mov. 14			
Passes to GMP Class A Yes/No		Acceptance Criteria : Penetration to be < 0.003 %						Label Attached	ves		
Counter Serial No Calibrated		Media <0.003 %	Media: <0.003 % Seals: <0.003 % Pass	Seals <0.003 %	Seals :			Customer Signature : Date: Next Test Due	 May-14		

The Validation Game: Check, Calibrate and Cascade

Closing the loop

- Facilities have had the equipment serviced
 - Where do the reports go?
 - What do the reports mean?
 - Actions to be taken?
 - Who needs to know the result and how is the information passed on?

Lay interpretation of engineers reports

HSE CL3 inspection question – How do you know a user understands the results of the inspection and the written report?

- Verbal report from engineer to user/facilities before leaving
 - Did it pass
 - Any concerns going forward
 - Close the loop!

Permits to work

- NOT just relevant to higher CL3 containment also CL2
- What information was the service engineer given?
- What was disinfected?
- What they can and cannot touch?
- What PPE should they wear?
- What should they do if the fire alarms go off?

The Validation Game: Check, Calibrate and Cascade

Record keeping - keeping evidence and being able to provide it for audit /inspection to HSE and others

- Permits and service reports
- Maintenance inspection schedules
 - Who is responsible
 - Who keeps the schedules
 - How do the users know service is required (before it becomes an issue)
 - Autoclaves annually under the written scheme (after that date CAN NOT BE USED)
 - Bureau Veritas SWIFT database
- A clear, formalized check list system to mange the schedule of inspections, servicing etc.
- A clear indication of who is responsible and who is responsible if that person is not available.
- A schedule to check the schedule Seriously!! Comment from HSE inspection
- RECORD, RECORD Time, Date and Sign